BMB Reports - Manuscript Submission

Manuscript Draft

Manuscript Number: BMB-22-030

Title: In Vivo Molecular and Single Cell Imaging

Article Type: Mini Review

Keywords: in vivo molecular imaging; optical imaging; MRI; PET; single cell

imaging

Corresponding Author: Kyung Oh Jung

Authors: Seongje Hong¹, Siyeon Rhee², Kyung Oh Jung^{1,3,*}

Institution: ¹Department of Anatomy, Chung-Ang University, ²Stanford Cardiovascular Institute and ³Department of Radiation Oncology, Stanford University,

Manuscript Type: Mini Review

Title: In Vivo Molecular and Single Cell Imaging

Author's name: Seongje Hong¹, Siyeon Rhee², Kyung Oh Jung^{1,3}

Affiliation:

- 1. Department of Anatomy, College of Medicine, Chung-Ang University, Seoul, Korea.
- 2. Stanford Cardiovascular Institute, Stanford University School of Medicine, Stanford,
- Department of Radiation Oncology, Stanford University School of Medicine, Stanford,
 CA, USA

Running Title: In vivo molecular imaging is useful for single cell imaging

Keywords: *in vivo* molecular imaging, optical imaging, MRI, PET, single cell imaging

Corresponding Author's Information: Kyung Oh Jung

Abstract

Molecular imaging is used to improve the disease diagnosis, prognosis, monitoring of treatment in living subjects. Numerous molecular targets have been developed for various cellular and molecular processes in genetic, metabolic, proteomic, and cellular biologic level. Molecular imaging modalities such as Optical Imaging, Magnetic Resonance Imaging (MRI), Positron Emission Tomography (PET), Single Photon Emission Computed Tomography (SPECT), and Computed Tomography (CT) can be used to visualize anatomic, genetic, biochemical, and physiologic changes in vivo. For in vivo cell imaging, certain cells such as cancer cells, immune cells, stem cells could be labeled by direct and indirect labeling methods to monitor cell migration, cell activity, and cell effects in cellbased therapy. In case of cancer, it could be used to investigate biological processes such as cancer metastasis and to analyze the drug treatment process. In addition, transplanted stem cells and immune cells in cell-based therapy could be visualized and tracked to confirm the fate, activity, and function of cells. In conventional molecular imaging, cells can be monitored in vivo in bulk non-invasively with optical imaging, MRI, PET, and SPECT imaging. However, single cell imaging in vivo has been a great challenge due to an extremely high sensitive detection of single cell. Recently, there has been great attention for in vivo single cell imaging due to the development of single cell study. In vivo single imaging could analyze the survival or death, movement direction, and characteristics of a single cell in live subjects. In this article, we reviewed basic principle of in vivo molecular imaging and introduced recent studies for in vivo single cell imaging based on the concept of in vivo molecular imaging.

INTRODUCTION

Molecular Imaging is a growing biomedical field focusing on the visualization, character ization, and quantification of biological processes in living subjects. Molecular imaging is used to improve the disease diagnosis, prognosis, monitoring of treatment in living subject s (1,2). Current research in molecular imaging require a multidisciplinary approach for var ious research fields such as molecular biology, chemistry, medical physics, engineering, biomedical imaging, and computer science. Various molecular images allow us to visu cellular and subcellular processes with genetic, metabolic, proteomic, and cellular biologic imaging in vitro, in vivo, and even 1, 2, ... M cular images such as Optical Imaging, Magnetic Resonance Imaging (MRI), Positron Emis sion Tomography (PET), Single Photon Emission Computed Tomography (SPECT), and C omputed Tomography (CT) could visualize anatomic, genetic, biochemical, and physiologi c changes (Figure 1). For molecular imaging, numerous molecular targets have been devel oped for cellular and molecular processes in genetic, metabolic, proteomic, and cellular bio logic level. As strategies for molecular imaging, there are direct and indirect labeling meth ods to monitor cells in small animal (Figure 2). Direct cell labeling methods usually use a t argetspecific probe which could be interacted with a specific target in cells. As molecular imagi ng probe, nanoparticles, antibodies, peptides and aptamers could be synthesized with other materials for imaging signals such as fluorescence dye and radioisotopes. For indirect label ing, reporter genebased imaging system in target cells could be used to monitor gene expression. The reporte r gene could encode detectable proteins based on the specific promoter, and the reporter

proteins with specific imaging signals could provide indirect information for reporter gene expression. Reporter gene imaging systems were used in optical, nuclear medicin e, and magnetic resonance imaging. Through these labeling methods, the ultimate goal of molecular imaging can diagnose diseases at the cellular level and noninvasively monitor the biochemical processes of cells in real time. In cell-based thera peutics, identifying migration pathways of cells are important issues and essential for i n vivo application (3,4). Various cell-based therapies using immune cells and stem cells f their location a n d biodistribution o c u s on vivo, their effects, and the presence or absence of toxicity in vivo In addition, suppression of cellcancer through based therapies requires monitoring and quantification of cells in vivo. In vivo cell imag ing through molecular imaging modalities could be useful tools for disease diagnosis a treatment monitoring methods for various diseases (3, 4, 7). n d In the process of treating cardiovascular diseases using stem cells, cell tracking could be used for the cell activity. For these reasons, in vivo cell imaging could be considered importantly for in vivo application. In terms of in vivo cell imaging, s ingle cell imaging and tracking have the greatest advantages to identify individual characteristics of each cell in the body (8-10). With reali time single cell m a g ing in vivo, specific characteristics of various single cells can be identified, and it could be o s s i b l e broaden the research direction based therapy in the future. This review describes various studies for in vivo molecular imaging tools with their advantages and disadvantages and we would like to re view the recent research related to single cell imaging and discuss the future directions of in vivo single cell imaging.

In vivo molecular imaging

Optical imaging

Fluorescence imaging and bioluminescence imaging are types of optical imaging that generates light from different sources (1,2). Fluorescence imaging uses a fluorescent (fluorophore) which is excited by an external light source. In bioluminescence imaging, luciferase enzyme converts luciferin substrate to oxyluciferin in the presence of oxygen, magnesium, and ATP, producing emitted visible light. Luciferase mainly used as a bioluminescence reporter gene includ in g Firefly luciferase (Fluc), Renilla luciferase (Rluc) and gaussia luciferase (Gluc). D uring imaging acquisition, fluorescence imaging required specific optical filters for both excitation and emission wavelengths. but bioluminescence imaging only required emissi Cooled charge coupled device (CCD) camera reduce s filter. o n thermal noise and increase sensitivity. Both fluorescence and bioluminescence are powerf ul imaging modalities' to visualize molecular processes in vitro and in vivo, because these modalities are simple, cheap, and convenient comparing to other imaging modalities (3,4). For fluorescence imaging, it doesn't require a substrate and has a short acquisition time tha t can acquire images within seconds (Table 1). However, fluorescence imaging requires ex ternal light source, causing phototoxicity and photobleaching. In addition, fluorescence im aging has some limitations for in vivo application due to autofluorescence and imaging dep th. Bioluminescence imaging have the advantages for high sensitivity and specificity, high signal-tonoise ratio, and no phototoxicity (Table 1). But bioluminescence imaging also has the limit ed imaging depth like fluorescence imaging and requires a substrate for imaging. In additio

n, disadvantages of optical images are the limited quantification for imagin

Optical imaging could be used for in vivo cell imaging. U

g, and poor spatial resolution due to scatter (11).

nlike various molecular imaging methods such as nuclear medicine imaging, fluorescence imaging uses fluorescent proteins without a reporter probe to monitor gene expression, c ellular localization, and protein-protein interaction (12). GFPs have been widely used to v isualize the specific protein distribution in cells, as well as to analyze the biological response of cells. Various types of engineered fluorescent proteins such as enhanced GF Ps (eGFPs), Red Fluorescent Proteins (RFPs), mCherry, and tdTomato have been developed to improve the fluorescence brightness and photostability (12). In addition, flu orescent proteins with long emission wavelengths such as NIR-I (700-1000 nm) or NIR-II (1,000-1,700 nm) have been recently developed for research and clinical studies. NIR fluorescenc visualize imaging advantage deepe has to tissue structure in vivo and to monitor surgical margins intraoperatively in clinical applicati on. (13-15). Some studies were conducted with fluorescence imaging to visualize cancer metastasis, cancer cell division, apoptosis, and cell cycle in vivo (16). By injecting various fluorescence-tagged tumor cells in the body, tumor metastases could be monitored in specific organs in the body (17,18). Other study showed that Cancer ste m-like cells (CSCs) and non-stem cancer cells (NSCCs) were labeled with each fluoro phore and injected into the subcutaneous and spleen of nude mice for in vivo cell ima ging (19). The survival rate, tumorigenesis, and metastasis of CSCs against chemical d rugs were studied in vivo. By expressing fluorescence proteins in HT-1080 fibrosarcom a cells, the nuclear-cytoplasmic ratio was visualized (20). For cell cycle study, th e method using the fluorescence ubiquitination cell cycle indicator (FUCCI) is used to observe changes in the expression of fluorescent proteins depending on the cell cycle (21,

22). Fluorescence imaging are expected to be continuously studied and developed to analyze cell differentiation, mobility, and characteristics for *in vivo* application.

Various studies for *in vivo* imaging have been conducted using bioluminescence imagi ng (1,2). Originally, in vivo bioluminescence imaging was developed in bacterial injection models (23). In this study, Salmonella typhimurium was marked with bioluminescence protein, which is a bacterial luciferase, and localization of bioluminescent Salmonella was detected in animals. For in vivo cell imaging, various kinds of cells are commonly labeled with bioluminescence proteins such as Fluc and Gluc (3,4). Bioluminescence imaging are widely used for in vivo whole-body imaging than fluorescence imaging due to high sensitivity and low background signals (Table 1). The study showed that doxorubicin-resistant breast cancer cell lines were tagged with the renilla luciferase (Rluc), and Natural Killer cells were also tagged with the enhanced firefl v luciferase (effluc) (24). In addition, the expression of specific genes in the brain have been studied using bioluminescence imaging (25). Other in vivo study was conducted to quantify the level of specific gene expression by expressing Alzheimer's-inducing tau mutations with luciferase (26). There was study proliferation, migration and differentiation on neural stem cells using bioluminescence imaging (27). Bioluminescence Resonance Energy Transfer (BRET) is a method to study protein-protein interaction in living cells, based on the non-radioactive energy transfer exhibited by luminescence donors (28). Fluorescence resonance energy transfer (FRET) is also used for protein protein interaction, but there are some limitations due to the emission wavelength and autofluorescence in FRET. On the other hand, BRET's luminescence donor could emit light through the action with the substrate without phototoxicity or autofluorescence. However, the low signal output of luminescence donors reduces the resolution of BRET, and various luminescence donors have been developed to overcome this limitation (29). For example, NanoLuc

(Nluc) improves the resolution of BRET with 150-fold higher signal intensity compared to Rluc or Fluc (30,31). It could be applicable for various fields such as protein protein interaction, gene regulation, protein stability and imaging through NanoBRET. Studies for transmembrane receptors are being conducted using BRET (32). Ligands that specifically bind to G-Protein Coupled Receptos (GPCRs) are analyzed using BRET as biosensors (33). In addition, the BRET-based GPCR constitutive sensor was investigated for conformational changes.

Magnetic Resonance Imaging (MRI)

MRI is noninvasive imaging modality that uses for disease diagnosis, detection, and treatment monitor ing (1,2). MRI detects changes in the rotational axis of protons in the water and uses strong magnets in the body to align protons (3,4). A radio frequency stimulates the the of equilibrium. protons protons spin When radio frequency is turned off, the MRI sensor detects the energy which is released from the d protons realign by the magnetic field. The amount of energy varies depending on the chemical properties of the molecule and the environment, and substances such as contrast agents and s uperparamagnetic iron oxide (SPIO) nanoparticles are used to increase signal intensity. MRI has the advantage of acquiring anatomical and physiological information with high spatial resolution (Table 1). However, MRI has a low sensitivity for imaging agent s and molecular reactions. In addition, it requires highly trained person and high costs for imaging acquisition. For in vivo studies, MRI imaging is used for cell monitoring in cell-based therapies. The reporter genes used in MRI contains a cellular receptor, an enzyme coding gene, and an endogenous reporter gene (34). Transferrin receptor is

a commonly used as MRI reporter gene expressed on the cell surface membrane of the transfected cells, resulting in decreased T2 signaling (34,35). Tyrosinase (TYR) is an enzyme that produces melanin to chelate paramagnetic ions, inducing a high MR signal (36). β-galactosidase is an enzyme coding gene to produce strong T1 contrast with a contrast agent such as EgadMe (37). Ferritin is an endogenous reporter gene that binds and stores iron (34, 38,39). For *in vivo* cell imaging, SPIO nanoparticles have been widely used in MRI (40-42). The stem cells labeled with SPIO were monitored by MRI for diagnosis and therapies. Recent study used SPIO for MRI imaging of hepatocellular carcinoma, showing excellent targeting and transfection ability for liver cancer (43).

Nuclear Medicine Imaging

Many radionuclides for nuclear medicine imaging have been used for a diagnosis and t reatment of patients (1,2). The principle of nuclear medicine imaging is that radiation emitted from radioactive materials is detected by external detectors (3,4). For **PET** imaging, d 0 tope is labeled with PET carriers, emitting a positively charged electron. This positron inte racts with the electron, causing two photons in opposite directions by an annihilation reaction. The PET scanner identifies the pair of these photons, and the com puter reconstructs the data through specific algorithm (44). PET has advantages which are highly sensitive and quantitative, detecting the molecu lar probes of picomolar concentration (Table 1). SPECT is an imaging modality for detecting and imaging radiation emitted from gamma-emitting isotopes (1,2). T he **SPECT** detector identifies gamma rays, and threedimensional images reconstructed. Because gammaare

emitting isotopes occur in all directions, collimators are used to achieve a specific range of resolution.

Nuclear medicine imaging provides anatomic and functional information invasively (3,4). Nuclear medicine imaging has several advantages for high sensitivity, high quantification, no attenuation problems, and easy translation from animal to human. Currently, there has b een many approved imaging agents for clinical use and developing various novel agent s in research fields. However, nuclear medicine imaging has several disadvantages of p oor spatial resolution and limited signal to noise ratio and it requires high operating costs and health risk by radiation exposure. To improve spatial resolution, animal s u c h imaging instruments f o r small micro-PET have been developed for the basic research. For in vivo cell imaging, reporter gene syste ms were used with nuclear medicine imaging. The herpes simplex virus type 1-thymi dine kinase (HSV1-TK) is a reporter gene commonly used for PET imaging (1,2). Th e HSV1-TK enzyme can phosphorylate the radiolabeled substrate such as pyrimidine n ucleoside and purine analogs. The study imaged the cytotoxic T lymphocytes by expre ssing the HSV1-TK reporter gene (45). Dopamine 2 receptor (D₂R) is used as a nucle ar medicine reporter gene to combine the Dopamine 2 receptor ligand such as ¹⁸F-fluo roethylspiperone (46). Sodium iodide symporter (NIS) is the cell membrane protein us ed to uptake radionuclides such as ¹²⁵I and ¹³¹I (47). Somatostatin receptor (SSTR) rep orter gene is a membrane receptors which uses PET tracers such as ⁶⁸Ga-DOTATOC for the diagnosis of neuroendocrine tumors and other various types of tumors (48). Est rogen receptor is a reporter gene for a estrogen receptor ligand (49). Human Norepine phrine transporter (NET) is used for imaging probes such as ¹¹C-ephedrine (50,51). Fo r in vivo cell imaging, ¹⁸F-fluorodeoxyglucose (FDG) is widely used for glucose metabolism, showing different uptake between cancer cells and normal cells

(52). Since radioisotopes such as ¹⁸F and ¹¹C has a short-half life, radioistopes such as ⁸⁹Zr and ⁶⁴Cu could be studied for long-term imaging (53). For SPECT imaging, this stu dy monitored cell death with radiolabeled C2Am in mouse lymphoma models and in human colorectal xenografts (54).

In vivo single cell image

In vivo single cell imaging could be studied to identify the migration, characteristics, differentiation of single cells and effects of drug treatment. The size of single cells was generally known from 1 to 100 um, depending on the cell types. Since the size of single cells is extremely small comparing to the body, it's very challenging to track a single cell in vivo (8-10). Although in vivo molecular imaging is useful tools to monitor the distribution of cells in the body, it is generally possible to image the cells in bulk non-invasively, not to individual single cells. Recently, there has been great attention for in vivo single cell imaging due to the development of single cell study (55). If it is possible to track a single cell in vivo, it could be helpful to investigate various studies with oncology, immunology, and stem cell biology. Here, we would like to review the recent research for in vivo single cell imaging (Figure 3).

Optical imaging

The reason that optical imaging has limitations *in vivo* is that the emission maximum of D-luciferin is 562 nm, showing a low access to deep tissue. To overcome these limitations, bioluminescence imaging using the near-infrared (NIR) has been developed. AkaLumin e is a luciferin analogue in which the aromatic structure of D-Luciferin was replaced with a benzothiazole moiety, generating NIR bioluminescence for sensitive deep-tissue

imaging. The emission maximum of AkaLumine is 675 nm. AkaLimine could react w ith Fluc emitting the NIR wavelength light. This methods could solve the limitations o f deep tissue in bioluminescence imaging for various animal study (56). The single cell imaging *in vivo* has been recently reported using the elevated AkaLumine, s howing higher signals in bioluminescence image and permeability to the bloodbrain barrier (BBB) than D-luciferin (57). this in vivo bioluminescence signals could be confirmed even in a single cell, detecting the sign als of cells in the lung of mice. For this experiment, the signal levels for Dluciferin/Fluc and AkaLumine/Akaluc in cells were compared in vivo. As a result, AkaLu mine/Akaluc combination showed ~52 fold stronger signals than D-luciferin/Fluc in vivo. Through these high signals, it was possible to visualize single cell in the body of mouse with optical imaging HeLa/Akaluc single cell injected d was into mouse a n the signals of a HeLa/Akaluc single cell was then confirmed in bioluminescence imagin results showed positive s i g n a 1 s in 2 out of 12 mice and it could be inferred that a single cell in 10 mice just passed through the lungs showing signals. I the experiments showed the stability of this molecule by penetrating the BBB on the brain. A single cell imaging with AkaLumine/Akaluc combination could allow realtime monitoring single cell. S i of n e AkaLumine could be used to monitor a single cell, it would be a promising approach for i *n vivo* cell tracking in the future.

Magnetic Resonance Imaging (MRI)

MRI has sensitivity for molecular probes in the range of micromolar to millimolar concent rations. Various studies for cell imaging have been performed by MRI (58). The showed that primary mouse hepatocytes with iron oxide particles and fluorescent agents and transplanted them into the mouse sple 5 9 S () i h en e epatocytes can migrate from the spleen to the liver as a single cell level, in vivo MRI could detect a black contrast area from a single cell in the livers. In addition, high-resolution MRI could be possible to image single cells in vivo by labeling cells with micron-size d iron oxide (MPIO) (60). In this study, MPIO could be taken up by a various cell ty pes and the best strategies to label cells, because it could uptake approximately ~3 fol d higher than nanometer-sized particles. Therefore, it could suggest various research d irections for in vivo single cell imaging using MRI. On the other hand, the study coul d monitor the metastasis of single breast cancer to the brain in vivo using MRI (61). I n this study, the MDA-MB-231BR cells labeled with MPIO were injected through the left ventricle of a heart in mouse. Over the time, the micro-metastasis of labeled cell s was visualized on the brain in single cell level using MRI, showing the tumor form ation on the brain from day 28 to day 33. These above research results showed that MRI could be a useful tools to monitor single cells and investigate the micro-metastas is of a single cancer cell in vivo. Therefore, the technology for diagnosing initial tumo r metastasis in vivo would be developed in single cell level and various research direc tions could be suggested in the future. However, MRI imaging have some limitations for in vivo real-time imaging. Since single cell imaging in in vivo MRI could be only possible at uniform-background anatomical organs such as brain and MRI doesn't ha ve the sufficient temporal resolution to track the moving cells in vivo, some improvem ent could be required in the future.

Intravital microscopy

Intravital microscopy (IVM) is a tool that can visualize several biological processes in live animals (62). With the development of intravital microscopy, biological research has been investigated to image in vivo subcellular structures. Single-cell pharmacokinetics imaging (SCPKI) could be used for the high resolution and temporal imaging of single cells in vivo (63). In this study, high-resolution imaging with intravital microscopy was used to measure drug kinetics through in vivo single cell imaging. A fluorescent derivative of a PARP inhibitor has been synthesized to measure drug distribution. The distribution of drugs could be visualized and analyzed over time according to the location of a single cell in vivo. In addition, it was possible to measure the proportion of cells which are r eceiving sub-therapeutic drugs or are not taking drugs. This quantitative data could be used to predict the effects of drugs in the body. Multiphoton intravital microscopy (MP-IVM) is a method that enables dynamic monitoring of cells under various physiological conditions in living animals (64). MP-IVM uses a longer wavelength (700 nm and greater) than confocal techniques to penetrate more deep tissue, and could be used to monitor the pharmacological action of drugs in vivo (65). The MP-IVM obtains high-contrast cellularlevel images from thick opaque specimens with minimized toxicity and photobleaching, and enables long-term imaging of biological processes in laboratory animals (66). MP-IVM was used to monitor leukocytes in vivo (67). In this study, the spatiotemporal dynamics of immune cells were visualized in mouse model of mandibular draining lymph nodes using MP-IVM. Through this, the pathogenesis of immune and inflammatory diseases in specific sites can be studied. Advances in these studies would allow pharmac okinetic and pharmacodynamic imaging to be used for the therapeutic efficacy in vivo. It also could have the potential on cell biology to investigate mechanisms for the

physiological and pathological characteristics. However, since IVM methods can analyze the characteristics of single cell on shallow tissues, not on the whole-body, it would be considered for *in vivo* whole-body experiments (68).

Photoacoustic imaging

Photoacoustic Т Α imaging is a molecular imaging tool that could image organs in a living subject with high contrast and high spatial resolution (69). PAT uses the photoacoustic effect, and the phot oacoustic energy is converted from absorbed light energy. PAT has the advantage for t he deep tissue penetration and high spatial resolution compared with pure optical imag ing (70). The photoacoustic imaging was investigated for in vivo studies related to Re (RBCs). As oxygen transporters, Blood C e 1 1 s RBCs important role in vivo, but the technology for i n imaging of single **RBCs** is lack currently. S ingle-RBC photoacoustic flowoxigraphy (FOG) was developed for i n study of oxygen metabolism in RBC (71). In this study, real-time spectral imaging of RBCs could be used in vivo. A single RBC FOG could image single erythrocyte oxyg enation on the brain of a mouse. In addition, this study showed correlation between singlestrong erythrocyte oxygenation and neurostimulation, and it was a novel approach to confirm the various through single-erythrocytes effects the brain. S ingle-RBC FOG has the advantage for not using a contrast agent which may affect the experime ntal results. This single-RBC FOG imaging modality could bring some benefits to investigate an early diagnosis an d treatment of tumors. With development of technologies, various studies related to tissue

oxygenation could be conducted in the future.

Nuclear Medicine Imaging

Among	molecular	imaging	modalities,	P	E	T			
has the highe	st sensitivity to d	etect the picomol	ar level of molec	cular targets.	It has b	een c			
onsidered as th	ne most promising	g method imaging	g modality for in	vivo single	cell trac	king.			
Originally, the concept of tracking in PET was used to measure the dynamics of powder an									
d fluid flows in chemical engineering (72). Then a mathematical frame was developed to tr									
acking the radioactive sources and radiolabeled cells (73). In addition, single cell tracking									
performance was validated in simulation and phantom studies using radioactive droplets (7									
4). Recent research showed in vivo real-time single cell tracking in whole-body animals									
with		PET	(1			
0). In this study, human breast cancer cells were labeled with mesoporous silica nanopartic									
les (MSNs) highly concentrating the ⁶⁸ Ga radioisotope. The labeled single cells were inject									
ed throu	gh the	tail vein	of m	nice, a	nd	real-			
time single	cell tracking	was performe	ed with PET	in vivo.	For	real-			
time single cell tracking, a novel single cell tracking algorithm was required with low temp									
oral resolution, because a temporal resolution of normal PET is several minutes to hour. In									
previous study, a novel trajectory reconstruction algorithm was developed for single cell tr									
acking, showing a low temporal resolution about 10 ms (73). To track a single cell in vivo t									
hrough this algorithm, the high radioactive single cells in which minimum radioactivity is									
> 20 Bq should be developed. Due to the superior properties such as a high surface area, M									
SNs could ferry radioisotopes efficiently. In conclusion, this PET imaging study revealed i									
n		vivo				real-			
time position of the injected single cells, inferring the movement speed and stationary site i									

n the body through imaging processing. This *in vivo* single cell tracking methods could be used to investigate the early stage of metastasis mechanism in cancer biology, and to study the kinetics of cell tracking on cell-based therapy such as regenerative therapy and imm unotherapy. According to the recent development of computer science, the imaging an alysis techniques were improved through imaging processing with deep learning in artific in a large in the large in the large in a dark in the continuous development of PET scanners and single cell biology, *in vivo* single cell tracking techniques would be very promising and powerful tools for research fields and colinical use.

ACKNOWLEDGMENTS

We thank to Suyeon Lee for the graphic illustration in Figure 2.

CONFLICTS OF INTEREST

The authors declare no conflict of interest.

REFERENCES

- Blasberg, R.G. and Gelovani-Tjuvajev, J. (2002) In vivo molecular-genetic Imaging. J Cell Biochem 39, 172-183
- Kang, J.H. and Chung, J.K. (2008) Molecular-genetic imaging based on reporter gene expression. J Nuc Med 49, 164S-179S
- 3. Kircher, M.F., Grambhir, S.S., Grimm, J. (2011) Noninvasive cell-tracking methods.

 Nat Rev Clin Oncol 8, 677-688
- 4. Youn, H. and Hong, K.J. (2012) In vivo Noninvasive Small Animal Molecular Imaging.

 Osong Public Health Res Perspect 3, 48-59
- 5. Sabapathy, V., Mentam, J., Jacob, P. M., Kumar, S. (2015) Noninvasive Optical Imaging and In Vivo Cell Tracking of Indocyanine Green Labeled Human Stem Cells Transplanted at Superficial Inor Depth Tissue of SCID Mice. Stem Cell 2015
- 6. Peñuelas, I., Mazzolini, G., Boán, J.F., et al (2005)

 Positron emission tomography imaging of adenoviralmediated transgene expression in liver cancer patients. Gastroenterolog 128, 1787-1795

- 7. Rogeers, W.J., Meyer, C.H., Kramer, C. (2006) Technology insight: in vivo cell tracking by use of MRI. Nat Clin Pract Cardiovasc Med 3, 554-562
- 8. Kwak, Y. H., Hong, S. M., & Park, S. S. (2010) A single cell tracking system in real-time. Cell Immuno. 265, 44-49
- 9. Carlson, A. L., Fujisaki, J., Wu, J., et al (2013) Tracking Single Cells in Live Animals
 Using a Photoconvertible Near-Infrared Cell Membrane Label. PLoS One 8, e69257
- 10. Jung, K.O., Kim, T.J., Yu, J.H., et al (2020) Whole-body tracking of single cells via positron emission tomography. Nat Biomed Eng 4, 835-844
- 11. Herschman, H.R. (2004). Noninvasive imaging of reporter gene expression in living subjects. Adv Cancer Res 92, 29-80
- 12. Shaner, N.C., Steinbach, P.A., Tsien, R.Y. (2005) A guide to choosing fluorescent proteins. Nature Methods 2, 905-909
- 13. Pan, K., Deng, H., Hu, S., et al (2020) Real-time surveillance of surgical margins via ICG-based near-infrared fluorescence imaging in patients with OSCC. World J Surg Oncol 18, 96
- 14. Chiu, C.H., Chao, Y.K., Liu, Y.H., et al (2016) Clinical use of near-infrared fluorescence imaging with indocyanine green in thoracic surgery: a literature review. J Thorac Dis 8, S744-S748
- 15. Kosaka, N., Ogawa, M., Choyke, P.L., Kobayashi, H. (2009) Clinical implications of n ear-infrared fluorescence imaging in cancer. Future oncology 5, 1501-1511
- 16. Hoffman, R.M. (2015) Application of GFP imaging in cancer. Lab Invest 95, 432-452
- 17. Yang, M., Baranov, E., Jiang, P., et al (2000) Whole-body optical imaging of green fluorescent protein-expressing tumors and metastases. Proceedings of the National Academy of Sciences 97, 1206-1211

- 18. Bouvet, M., Wang, J., Nardin, S.R., et al (2002) Real-Time Optical Imaging of Primary Tumor Growth and Multiple Metastatic Events in a Pancreatic Cancer Orthotopic Model. Cancer Res 62, 1534-1540
- 19. Suetsugu, A., Osawa, Y., Nagaki, M., et al (2010) Simultaneous color-coded imaging to distinguish cancer "stem-like" and non-stem cells in the same tumor. J Cell Biochem 111, 1035-1341
- 20. Suetsugu, A., Jiang, P., Moriwaki, H., Saji, S., Bouvet, M., Hoffman, R.M. (2013)
 Imaging nuclear cytoplasm dynamics of cancer cells in the intravascular niche of live
 mice. Anticancer Res 33, 4229-4236
- 21. Yano, S., Zhang, Y., Zhao, M., et al (2014) Tumor-targeting Salmonella typhimurium A1-R decoys quiescent cancer cells to cycle as visualized by FUCCI imaging and become sensitive to chemotherapy. Cell Cycle 13, 3958-3963
- 22. Yano, S., Miwa, S., Mii, S., et al (2014) Invading cancer cells are predominantly in G0/G1 resulting in chemoresistance demonstrated by real-time FUCCI imaging. Cell Cycle 13, 953-960
- 23. Contag, C.H., Contag, P.R., Mullins, J.I., Spilman, S.D., Stevenson, D.K., Benaron, D.A. (1995) Photonic detection of bacterial pathogens in living hosts. Mol Microbiol 18, 593-603
- 24. Hwang, M.H., Li, X.J., Kim, J.E., et al (2015) Potential Therapeutic Effect of Natural Killer Cells on Doxorubicin-Resistant Breast Cancer Cells In Vitro. PLoS. One 10, e0136209
- 25. Hochgräfe, K. and Mandelkow, E.M. (2013) Making the Brain Glow: In Vivo Bioluminescence Imaging to Study Neurodegeneration. Mol. Neurobiol 47, 868-882
- 26. Van der Jeugd, A., Hochgräfe, K., Ahmed, T., et al (2012) Cognitive defects are reversible in inducible mice expressing pro-aggregant full-length human Tau. Acta

- Neuropathol 123, 787-805
- 27. Reumers, V., Deroose, C.M., Krylyshkina, O., et al (2008) Noninvasive and Quantitative Monitoring of Adult Neuronal Stem Cell Migration in Mouse Brain Using Bioluminescence Imaging. Stem cells 26, 2382-2390
- 28. Prinz, A., Diskar, M., & Herberg, F. W. (2006). Application of bioluminescence resonance energy transfer (BRET) for biomolecular interaction studies. ChemBioChem, 7, 1007-1012.
- 29. Kobayashi, H., Picard, L. P., Schönegge, A. M., & Bouvier, M. (2019).

 Bioluminescence resonance energy transfer—based imaging of protein—protein interactions in living cells. Nature protocols, 14, 1084-1107.
- 30. Dale, N. C., Johnstone, E. K., White, C. W., & Pfleger, K. D. (2019). NanoBRET: the bright future of proximity-based assays. Frontiers in Bioengineering and Biotechnology, 7, 56.
- 31. England, C. G., Ehlerding, E. B., & Cai, W. (2016). NanoLuc: a small luciferase is brightening up the field of bioluminescence. Bioconjugate chemistry, 27, 1175-1187.
- 32. El Khamlichi, C., Reverchon-Assadi, F., Hervouet-Coste, N., Blot, L., Reiter, E., & Morisset-Lopez, S. (2019). Bioluminescence resonance energy transfer as a method to study protein-protein interactions: application to G protein coupled receptor biology. Molecules, 24, 537.
- 33. Picard, L. P., Schönegge, A. M., Lohse, M. J., & Bouvier, M. (2018). Bioluminescence resonance energy transfer-based biosensors allow monitoring of ligand-and transducer-mediated GPCR conformational changes. Communications biology, 1, 1-7.
- 34. Yang, C., Tian, R., Liu, T., Liu, G. (2016) MRI Reporter Genes for Noninvasive Molecular Imaging. Molecules 21, 580
- 35. Zhang, R., Feng, G., Zhang, C.J., Cai, X., Cheng, X., Liu, B. (2016) Real-time specific

- light-up sensing of transferrin receptor (TfR): Image-guided photodynamic ablation of cancer cells through controlled cytomembrane disintegration. Anal. Chem 88, 4841-4848
- 36. Qin, C., Cheng, K., Chen, K., et al (2013) Tyrosinase as a multifunctional reporter gene for Photoacoustic/MRI/PET triple modality molecular imaging. Sci Rep 3, 1-8
- 37. Gilad, A. A., Winnard Jr, P. T., van Zijl, P. C., & Bulte, J. W. (2007) Developing MR reporter genes: promises and pitfalls. NMR Biomed 20, 275-290
- 38. Vande Velde, G., Rangarajan, J.R., Toelen, J., et al (20 1 1)

 Evaluation of the specificity and sensitivity of ferritin as an MRI reporter gene in the mo
 use brain using lentiviral and adeno-associated viral vectors. Gene Ther 18, 594-605
- 39. K i m , H . S . , W o o , J . , L e e , J . H ., e t a l (2015)

 In vivo Tracking of Dendritic Cell using MRI Reporter Gene, Ferritin. PLoS One 10, e0125291
- 40. Gilad, A.A., McMahon, M.T., Walczak, P., e t a 1 (2 0 0 7)

 Artificial reporter gene providing MRI contrast based on proton exchange. Nat Biotech nol 25, 217-219
- 41. Massoud, T.F. a n d Gambhir, S.S. (2003)

 Molecular imaging in living subjects: seeing fundamental biological processes in a new light. Genes Dev 17, 545-580
- 42. Bull, E., Madani, S.Y., Sheth, R., Seifalian, A., Green, M., Seifalian, A.M. (2014)

 Stem cell tracking using iron oxide nanoparticles. Int J Nanomedicine 9, 1641
- 43. Fan, K., Lu, C., Shu, G., et al (2021) Sialic acid-engineered mesoporous polydopamine dual loaded with ferritin gene and SPIO for achieving endogenous and exogenous synergistic T2-weighted magnetic resonance imaging of HCC. J Nanobiotec hnology 19, 1-17

- 44. Omami, G., Tamimi, D., Branstetter, B.F., et al 18F-FDG-(2014)Basic principles and applications of PET/CT in oral and maxillofacial imaging: A pictorial essay. Imaging Sci Dent 44, 325-332
- 45. Keu, V.K., Witney, T.H., Yaghoubi, S., et al (2017) Reporter gene imaging of targ eted T cell immunotherapy in recurrent glioma. Sci Transl Med 9, eaag2196
- 46. MacLaren, D.C., Gambhir, S.S., Satyamurthy, N., et al (1999) Repetitive, non-invasive imaging of the dopamine D2 receptor as a reporter gene in living animals.

 Gene therapy 6, 785-791
- 47. Chung, J.K. (2002) Sodium iodide symporter: its role in nuclear medicine. Journal of Nuclear Medicine 43, 1188-1200
- 48. Rinne, P., Hellberg, S., Kiugel, M., et al (2016)

 Comparison of Somatostatin Receptor 2
 Targeting PET Tracers in the Detection of Mouse Atherosclerotic Plaques.

 Molecular Imaging and Biology 18, 99-108
- 49. Furukawa, T., Lohith, T. G., Takamatsu, S., Mori, T., Tanaka, T., & Fujibayashi, Y. (2 0 0 6) Potential of the FES-hERL PET reporter gene system—Basic evaluation for gene therapy monitoring. Nuclear medicine and biology 33, 145-1 51
- 50. Alt mann, A., Kissel, M., Zitzmann, S., et al (2003) Increased MIBG uptake after transfer of the human norepinephrine transporter gene in r at hepatoma. Journal of Nuclear Medicine 44, 973-980
- 51. Moroz, M.A., Serganova, I., Zanzonico, P., et al (2007)

 Imaging hNET reporter gene expression with 124I-MIBG. Journal of Nuclear Medicine
 48, 827-836
- 52. Berberat, P., Friess, H., Kashiwagi, M., Beger, H. G., & Büchler, M. W (1999)

 Diagnosis and Staging of Pancreatic Cancer by Positron Emission Tomography. World

- journal of surgery 23, 882-887
- 53. Stockhofe, K., Postema, J. M., Schieferstein, H., & Ross, T. L..(2014) Radiolabeling of Nanoparticles and Polymers for PET Imaging. Pharmaceuticals7, 392-418
- 54. Neves, A.A., Xie, B., Fawcett, S., et al (2017)

 Rapid imaging of tumor cell death in vivo using the C2A domain of Synaptotagmin-I. J

 Nucl Med 58, 881-887
- 55. Schier, A.F. (2020) Single-cell biology: beyond the sum of its parts. Nat Methods 17, 17–20
- 56. Kichimaru, T., Iwano, S., Kiyama, M., et al (2016)

 A luciferin analogue generating nearinfrared bioluminescence achieves highly sensitive deep-tissue imaging.

 Nature communications 7, 1-8
- 57. Iwano, S., Sugiyama, M., Hama, H., et al (2018) Single-cell bioluminescence imaging of deep tissue in freely moving animals. Science 359, 935
- 58. Shapiro, E. M., Skrtic, S., Sharer, K., Hill, J. M., Dunbar, C. E., & Koretsky, A. P. (2004) MRI detection of single particles for cellular imaging. Proc Natl Acad Sci U S A 101, 10901-10906
- 59. Shapiro, E. M., Sharer, K., Skrtic, S., & Koretsky, A. P. (2006) In vivo detection of single cells by MRI. Magn Reson Med 55, 242-249
- 60. Shapiro, E. M., Skrtic, S., & Koretsky, A. P. (2 0 0 5)

 Sizing it up: Cellular MRI using micron-sized iron oxide particles. Magn Reson Med 5
 3, 329-338
- 61. Heyn, C., Ronald, J.A., Ramadan, S.S., et al (2006)

 In vivo MRI of cancer cell fate at the singlecell level in a mouse model of breast cancer metastasis to the brain. Magn Reson Med
 56, 1001-1010

- 62. Masedunskas, A., Milberg, O., Porat-Shliom, N., Sramkova, M., Wigand, T., Amornphimoltham, P., Weigert, R. (2012)

 Intravital microscopy: a practical guide on imaging intracellular structures in live animal s. Bioarchitecture 2, 143-157
- 63. Thur b er, G.M., Y a n g, K.S., R e i n e r, T., et al (2013) Single-cell and subcellular pharmacokinetic imaging allows insight into drug action in vivo. Na t Commun 4, 1-10
- 64. Hoover, E. E., & Squier, J. A. (2013). Advances in multiphoton microscopy technology. Nature photonics, 7, 93-101.
- 65. Morimoto, A., Kikuta, J., & Ishii, M. (2020). Intravital multiphoton microscopy as a novel tool in the field of immunopharmacology. Pharmacology & Therapeutics, 206, 107429.
- 66. Liang, Y., & Walczak, P. (2021). Long term intravital single cell tracking under multiphoton microscopy. Journal of Neuroscience Methods, 349, 109042.
- 67. Lopez, M. J., Seyed-Razavi, Y., Yamaguchi, T., Ortiz, G., Sendra, V. G., Harris, D. L., ... & Hamrah, P. (2020). Multiphoton intravital microscopy of mandibular draining lymph nodes: a mouse model to study corneal immune responses. Frontiers in Immunology, 11, 39.
- 68. Carney, B., Kossatz, S., & Reiner, T. (2017) Molecular Imaging of PARP.

 J Nucl Med 58, 1025-1030
- 69. Xu, M. and Wang, L.V. (2006) Photoacoustic imaging in biomedicine. Review of scientific instruments 77, 041101
- 70. Xia, J., Yao, J., & Wang, L. V. (2014)

 Photoacoustic tomography: principles and advances. Electromagnetic waves 147, 1-22
- 71. Wang, L., Maslov, K., & Wang, L. V. (2013) Single-cell label-free photoacoustic flowoxigraphy in vivo. Proc Natl Acad Sci 110, 5759-5764

- 72. Parker, D., Broadbent, C., Fowles, P., Hawkesworth, M., McNeil, P. (1993) Positron emission particle tracking a technique for studying fow within engineering equipment.

 Nucl. Instrum. Methods Phys. Res. A 326, 592–607
- 73. Lee, K., Kim, T.J., Pratx, G. (2015) Single-cell tracking with PET using a novel trajectory reconstruction algorithm. IEEE Trans. Med. Imaging 34, 994–1003
- 74. Ouyang, Y., Kim, T.J., Pratx, G. (2016) Evaluation of a BGO-based PET system for single-cell tracking performance by simulation and phantom studies. Mol. Imaging 15, 1–8




Figure 1. Molecular imaging modalities for in vivo cell imaging




Figure 2. Labeling methods for in vivo cell imaging






Figure 3. Studies for in vivo single cell imaging and tracking

TABLE

lmaging Modalities	Penetration	Sensitivity (mol/L)	Resolution (mm)	Time	
Fluorescence	Poor	10 ⁻⁹ ~10 ⁻¹²	0.2-2 mm	Sec~Min	
Bioluminescence	Fair	$10^{-15} \sim 10^{-17}$	0.2-2 mm	Min	
PET	Good	$10^{-11} \sim 10^{-12}$	1-2 mm	Min	
SPECT	Good	$10^{-10} \sim 10^{-11}$	1-2 mm	Min	
MRI	Excellent	$10^{-3} \sim 10^{-5}$	10-100 um	Min~Hr	

Table 1. The properties of molecular imaging modalities

In vivo cell imaging

Cancer cells

Stem cells

Immune cells

:

Other types of cells

Genetic Change - Molecular-genetic imaging


Biochemical Change - PET, MRS

Physiological Change - SPECT, MRI


Anatomical Change - CT, MRI

Direct Labeling

Indirect Labeling



Intravital microscopy imaging




Using a fluorescent derivative PARP inhibitor

A single cancer cell image in IVM, showing drug kinetics through in vivo single cell imaging

Thurber, G. M. et al. in 2013

Bioluminescence imaging




Iwano, S. et al. in 2018

- ✓ Using AkaLumine/Akaluc combination
- A single cancer cell image in bioluminescence imaging, showing higher signals and permeability to the blood-brain barrier



Photoacoustic imaging




Using Single-RBC photoacoustic flowoxigraphy


A single red blood cell image in PAT, showing single erythrocyte oxygenation in the brain of a mouse

Wang, L. et al. in 2013

MRI imaging




Heyn, C. et al. in 2006



A single cancer cell image in MRI, showing the micro-metastasis of a single cancer on the brain

PET imaging



Jung, K. O. et al. in 2020

- Using 68Ga labeled mesoporous silica nanoparticles (MSNs)
 - A single cancer cell image in PET, showing in vivo real-time location of injected single cell in whole-body